LIFEPAK® 1000 DEFIBRILLATOR
Battery Removal and Reinstallation CHECKLIST

Unit Serial Number_____
Department/Location_____

If your device powers off unexpectedly, either during inspection or during patient treatment, immediately remove and reinstall your existing battery to restore power to the device. If power is not restored, replace the battery with a spare battery and call Physio-Control immediately to arrange for servicing of your device at 1-866-231-1220, 6:00 A.M. to 4:00 P.M. (Pacific), Monday – Friday.

Instruction		
Remove battery and reinstall weekly. The removal and reinstallation of the battery will clean the contacts of oxidation and will reduce the likelihood of oxidation occurring.		
Battery removed and reinstalled	**Date**	**Initials**
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		
☐		