

LIFEPAK[®] 15 monitor/defibrillator

Data Sheet

Ease of use

- More than 20 clearly labeled, dedicated buttons
- Front-facing cables, connectors, and 100mm printer

Clinical effectiveness

- 360J biphasic
- ST-segment trending
- University of Glasgow 12-Lead ECG Analysis Program

Durability

- LIFEPAK TOUGH[™]
- Dual-layer screen protector
- Large, shock absorbing handle

Reliability

- Upgradeable platform
- Daily, 3AM diagnostic self-test



Specifications

General

The LIFEPAK 15 monitor/defibrillator has six main operating modes:

AED mode: for automated ECG analysis and a prompted treatment protocol for patients in cardiac arrest.

Manual mode: for performing manual defibrillation, synchronized cardioversion, noninvasive pacing, and ECG and vital sign monitoring.

Archive mode: for accessing stored patient information.

Setup mode: for changing default settings of the operating functions.

Service mode: for authorized personnel to perform diagnostic tests and calibrations.

Demo mode: for simulated waveforms and trend graphs for demonstration purposes.

Physical characteristics

Weight:

- Basic monitor/defibrillator with new roll paper and two batteries installed: 7.9 kg (17.5 lb)
- Fully featured monitor/defibrillator with new roll paper and two batteries installed: 8.4 kg (18.5 lb)

Lithium-ion battery: ≤ 0.6 kg (1.3 lb)

Accessory bags and shoulder strap: 1.77 kg (3.9 lb)

Standard (hard) paddles: 0.95 kg (2.1 lb)

Height: 31.7 cm (12.5 in)

Width: 40.1 cm (15.8 in)

Depth: 23.1 cm (9.1 in)

Display

Size (active viewing area): 212 mm (8.4 in) diagonal; 171 mm (6.7 in) wide x 128 mm (5.0 in) high

Resolution: display type 640 dot x 480 dot color backlit LCD

User selectable display mode: full color or SunVue[™] display high contrast

Display: a minimum of 5 seconds of ECG and alphanumerics for values, device instructions, or prompts

Display: up to three waveforms

Waveform display sweep speed: 25 mm/sec for ECG, SpO₂, IP, and 12.5 mm/sec for CO₂

Data management

The device captures and stores patient data, events (including waveforms and annotations), and continuous waveform and patient impedance records in internal memory.

The user can select and print reports, and transfer the stored information via supported communication methods.

Report types:

- Three format types of CODE SUMMARY[™] critical event record: short, medium, and long
- 12-lead ECG with STEMI statements
- Continuous Waveform (transfer only)
- Trend Summary
- Vital Sign Summary
- Snapshot

Memory capacity: Total capacity is 360 minutes of continuous ECG, 90 minutes of continuous data from all channels, or 400 single waveform events. Maximum memory capacity for a single patient includes up to 200 single waveform reports and 90 minutes of continuous ECG.

Communications

The device is capable of transferring data records by wired or wireless connection. This device complies with Part 15 of the FCC rules, and its operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any

interference received, including interference that may cause undesired operation.

- Serial Port RS232 communication + 12V available
- Limited to devices drawing maximum 0.5 A current
- Bluetooth[®] technology provides short-range wireless communication with other Bluetooth-enabled devices

Monitor

ECG

ECG is monitored via several cable arrangements:

A 3-wire cable is used for 3-lead ECG monitoring.

A 5-wire cable is used for 7-lead ECG monitoring.

A 10-wire cable is used for 12-lead ECG acquisition. When the chest electrodes are removed, the 10-wire cable functions as a 4-wire cable.

Standard paddles or QUIK-COMBO pacing/defibrillation/ECG electrodes are used for paddles lead monitoring.

Frequency response:

- Monitor: 0.5 to 40 Hz or 1 to 30 Hz
- Paddles: 2.5 to 30 Hz
- 12-lead ECG diagnostic: 0.05 to 150 Hz

Lead selection:

- Leads I, II, III, (3-wire ECG cable)
- Leads I, II, III, AVR, AVL, and AVF acquired simultaneously (4-wire ECG cable)
- Leads I, II, III, AVR, AVL, AVF, and C lead acquired simultaneously (5-wire ECG cable)
- Leads I, II, III, AVR, AVL, AVF, V1, V2, V3, V4, V5, and V6 acquired simultaneously (10-wire ECG cable)

ECG size: 4, 3, 2.5, 2, 1.5, 1, 0.5, 0.25 cm/mV (fixed at 1 cm/mV for 12-lead)

Specifications cont.

Heart rate display:

- 20–300 bpm digital display
- Accuracy: $\pm 4\%$ or ± 3 bpm, whichever is greater
- QRS Detection Range Duration: 40 to 120 msec
- Amplitude: 0.5 to 5.0 m

Common mode rejection (CMRR): ECG Leads: 90 dB at 50/60 Hz

SpO₂/SpCO/SpMet

Sensors:

- MASIMO® sensors including RAINBOW® sensors
- NELLCOR® sensors when used with the MASIMO RED™ MNC adapter

SpO₂

Displayed saturation range: “<50” for levels below 50%; 50 to 100%

Saturation accuracy: 70–100% (0–69% unspecified)

Adults/pediatrics:

± 2 digits (during no motion conditions)

± 3 digits (during motion conditions)

Dynamic signal strength bar graph

Pulse tone as SpO₂ pulsations are detected

SpO₂ update averaging rate user selectable:

4, 8, 12 or 16 seconds

SpO₂ sensitivity user selectable: Normal, High

SpO₂ measurement: Functional SpO₂ values are displayed and stored

Pulse rate range: 25 to 240 bpm

Pulse rate accuracy (adults/pediatrics):

± 3 digits (during no motion conditions)

± 5 digits (during motion conditions)

Optional SpO₂ waveform display with autogain control

SpCO®

SpCO concentration display range: 0 to 40%

SpCO accuracy: ± 3 digits

SpMET®

SpMet saturation range: 0 to 15.0%

SpMet display resolution: 0.1% up to 10%

SpMet accuracy: ± 1 digit

NIBP

Blood pressure systolic pressure range: 30 to 255 mmHg

Diastolic pressure range: 15 to 220 mmHg

Mean arterial pressure range: 20 to 235 mmHg

Units: mmHg

Blood pressure accuracy: ± 5 mmHg

Blood pressure measurement time: 20 seconds, typical (excluding cuff inflation time)

Pulse rate range: 30 to 240 pulses per minute

Pulse rate accuracy: ± 2 pulses per minute or $\pm 2\%$, whichever is greater

Operation features initial cuff pressure: User selectable, 80 to 180 mmHg

Automatic measurement time interval: User selectable, from 2 min to 60 min

Automatic cuff deflation excessive pressure: If cuff pressure exceeds 290 mmHg

Excessive time: If measurement time exceeds 120 seconds

CO₂

CO₂ range: 0 to 99 mmHg (0 to 13.2 kPa)

Units: mmHg, %, or kPa

Respiration rate accuracy:

- 0 to 70 bpm: ± 1 bpm
- 71 to 99 bpm: ± 2 bpm

Respiration rate range: 0 to 99 breaths/minute

Rise time: 190 msec

Response time: 4.3 seconds (includes delay time and rise time)

Initialization time: 30 seconds (typical), 10–180 seconds

Ambient pressure: automatically compensated internally

Optional display: CO₂ pressure waveform

- Scale factors: Autoscale, 0–20 mmHg (0–4 Vol%), 0–50 mmHg (0–7 Vol%), 0–100 mmHg (0–14 Vol%)

Invasive pressure

Transducer type: Strain-gauge resistive bridge

Transducer Sensitivity: 5 μ V/V/mmHg

Excitation voltage: 5 Vdc

Connector: Electro Shield: CXS 3102A 14S-6S

Bandwidth: Digital filtered, DC to 30 Hz (< -3db)

Zero drift: 1 mmHg/hr without transducer drift

Zero adjustment: ± 150 mmHg including transducer offset

Numeric accuracy: ± 1 mmHg or 2% of reading, whichever is greater, plus transducer error

Pressure range: -30 to 300 mmHg, in six user selectable ranges

Invasive pressure display

Display: IP waveform and numerics

Units: mmHg

Labels: P1 or P2, ART, PA, CVP, ICP, LAP (user selectable)

Temperature

Range: 24.8° to 45.2°C (76.6° to 113.4°F)

Resolution: 0.1°C

Accuracy: $\pm 0.2^\circ\text{C}$ including sensor

Reusable temperature cable: 5 foot or 10 foot

Disposable sensor types: Surface–Skin; Esophageal/Rectal

Trend

Time scale: Auto, 30 minutes, 1, 2, 4, or 8 hours

Duration: Up to 8 hours

ST segment: After initial 12-lead ECG analysis, automatically selects and trends ECG lead with the greatest ST displacement

Display choice of: HR, PR (SpO₂), PR (NIBP), SpO₂ (%), SpCO (%), SpMet (%), CO₂ (EtCO₂/FiCO₂), RR (CO₂), NIBP, IP1, IP2, ST

Alarms

Quick set: Activates alarms for all active vital signs

VF/VT alarm: Activates continuous (CPSS) monitoring in Manual mode

No breath alarm: Occurs when 30 seconds has elapsed since last detected respiration

Heart rate alarm limit range: Upper, 100–250 bpm; lower, 30–150 bpm

Interpretive algorithm

12-Lead interpretive algorithm: University of Glasgow 12-Lead ECG Analysis Program, includes AMI and STEMI statements

Printer

Prints continuous strip of the displayed patient information and reports

Paper size: 3.9 in (100 mm)

Print speed: 25 mm/sec or 12.5 mm/sec

- Optional: 50 mm/sec time base for 12-lead ECG reports

Delay: 8 seconds

Autoprint: Waveform events print automatically

Frequency response:

- Diagnostic: 0.05 to 150 Hz or 0.05 to 40 Hz
- Monitor: 0.67 to 40 Hz or 1 to 30 Hz

Defibrillator

Biphasic waveform: Biphasic Truncated Exponential

The following specifications apply from 25 to 200 ohms, unless otherwise specified:

Energy accuracy: ± 1 joule or 10% of setting, whichever is greater, into 50 ohms, ± 2 joules or 15% of setting, whichever is greater, into 25–175 ohms.

Voltage compensation: Active when disposable therapy electrodes are attached. Energy output within $\pm 5\%$ or ± 1 joule, whichever is greater, of 50 ohms value, limited to the available energy which results in the delivery of 360 joules into 50 ohms.

Paddle options: QUIK-COMBO® pacing/defibrillation/ECG electrodes (standard). Cable Length 8 foot long (2.4 m) QUIK-COMBO cable (not including electrode assembly).

Standard paddles (optional)

Manual mode

Energy select: 2, 3, 4, 5, 6, 7, 8, 9, 10, 15, 20, 30, 50, 70, 100, 125, 150, 175, 200, 225, 250, 275, 300, 325, and 360 joules

Charge time: Charge time to 360 joules in less than 10 seconds, typical

Synchronous cardioversion: Energy transfer begins within 60 msec of the QRS peak

Paddles leads off sensing: When using QUIK-COMBO electrodes, the device indicates Paddles Leads Off if the resistive part of the patient impedance is greater than 300 $\pm 15\%$ ohms, or if the magnitude of the patient impedance is greater than 440 $\pm 15\%$ ohms.

AED Mode

Shock Advisory System™ (SAS): an ECG analysis system that advises the operator if the algorithm detects a shockable or non-shockable ECG rhythm. SAS acquires ECG via therapy electrodes only.

Shock ready time: Using a fully charged battery at normal room temperature, the device is ready to shock within 20 seconds if the initial rhythm finding is “SHOCK ADVISED”

Biphasic output: Energy Shock levels ranging from 150–360 joules with same or greater energy level for each successive shock

cprMAX™ Technology: In AED mode, cprMAX™ technology provides a method of maximizing the CPR time that a patient receives, with the overall goal of improving the rate of survival of patients treated with AEDs.

Setup options:

- Auto Analyze: Allows for auto analysis. Options are OFF, AFTER 1ST SHOCK
- Initial CPR: Allows the user to be prompted for CPR for a period of time prior to other activity. Options are OFF, ANALYZE FIRST, CPR FIRST
- Initial CPR Time: Time interval for Initial CPR. Options are 15, 30, 45, 60, 90, 120, and 180 seconds.
- Pre-Shock CPR: Allows the user to be prompted for CPR while the device is charging. Options are OFF, 15, 30 seconds.
- Pulse Check: Allows the user to be prompted for a pulse check at various times. Options are ALWAYS, AFTER EVERY SECOND NSA, AFTER EVERY NSA, NEVER

- Stacked Shocks: Allows for CPR after 3 consecutive shocks or after a single shock. Options are OFF, ON
- CPR Time: 1 or 2 User selectable times for CPR. Options are 15, 30, 45, 60, 90, 120, 180 seconds and 30 minutes.

Pacer

Pacing mode: Demand or non-demand rate and current defaults

Pacing rate: 40 to 170 PPM

Rate accuracy: ±1.5% over entire range

Output waveform: Monophasic, truncated exponential current pulse (20 ± 1 ms)

Output current: 0 to 200 mA

Pulse: Pacing pulse frequency reduced by a factor of 4 when activated

Refractory period: 180 to 270 msec (function of rate)

Environmental

Unit meets functional requirements during exposure to the following environments unless otherwise stated.

Operating temperature: 0° to 45°C (32° to 113°F); -20°C (-4°F) for 1 hour after storage at room temperature; 60°C (140°F) for 1 hour after storage at room temperature

Storage temperature: -20° to 65°C (-4° to 149°F) except therapy electrodes and batteries

Relative humidity, operating: 5 to 95%, non-condensing. NIBP: 15 to 95%, non-condensing

Relative humidity, storage: 10 to 95%, non-condensing

Atmospheric pressure, operating: -382 to 4,572 m (-1,253 to 15,000 ft). NIBP: -152 to 3,048 m (-500 to 10,000 ft)

Water resistance, operating: IP44 (dust and splash resistance) per IEC 529 and EN 1789 (without accessories except for 12-lead ECG cable, hard paddles, and battery pack)

Vibration: MIL-STD-810E Method 514.4, Propeller Aircraft - category 4 (figure 514.4-7 spectrum a), Helicopter - category 6 (3.75 Grms), Ground Mobile - category 8 (3.14 Grms), EN 1789: Sinusoidal Sweep, 1 octave/min, 10-150 Hz, ±0.15 mm/2 g

Shock (drop): 5 drops on each side from 18 inches onto a steel surface EN 1789: 30-inch drop onto each of 6 surfaces

Shock (functional): Meets IEC 60068-2-27 and MIL-STD-810E shock requirements 3 shocks per face at 40 g, 6 ms half-sine pulses

Bump: 1000 bumps at 15 g with pulse duration of 6 msec

Impact, non-operating: EN 60601-1 0.5 + 0.05 joule impact UL 60601-1 6.78 Nm impact with 2-inch diameter steel ball. Meets IEC62262 protection level IK 04.

EMC: EN 60601-1-2:2006 Medical Equipment -General Requirements for Safety - Collateral

Standard: Electromagnetic Compatibility - Requirements and Tests EN 60601-2-4:2003: (Clause 36) Particular Requirements for the Safety of Cardiac Defibrillators and Cardiac Defibrillator-Monitors

Cleaning: Cleaning 20 times with the following: Quaternary ammonium, isopropyl alcohol, hydrogen peroxide

Chemical resistance: 60 hour exposure to specified chemicals: Betadine (10% Povidone-Iodine solution), Coffee, Cola, Dextrose (5% Glucose solution), Electrode Gel/Paste (98% water, 2% Carbopol 940), HCL (0.5% solution, pH=1), Isopropyl Alcohol, NaCl solution (0.9% solution), Cosmetic discoloration of the paddle well shorting bar shall be allowed following exposure to HCL (0.5% solution).

Power

Power adapters: AC or DC

Power Adapters provide operation and battery charging from external AC or DC power

- Full functionality with or without batteries when connected to external AC/DC
- Typical battery charge time while installed in LIFEPAK 15 device is 190 minutes
- Indicators: external power indicator, battery charging indicator

Dual battery: Capability with automatic switching

Low battery indication and message: Low battery fuel gauge indication and low battery message in status area for each battery

Replace battery indication and message: Replace battery fuel gauge indication, audio tones and replace battery message in the status area for each battery. When replace battery is indicated, device auto-switches to second battery. When both batteries reach replace battery condition, a voice prompt instructs user to replace battery.

Battery capacity

For two, new fully-charged batteries, 68°F (20°C)

Operating mode		Monitoring Pacing		Defibrillation
		(minutes)	(minutes)	(360J discharges)
Total capacity to shutdown	Typical	360	340	420
	Minimum	340	320	400
Capacity after low battery	Typical	21	20	30
	Minimum	12	10	6

Battery

Battery specifications

Battery type: Lithium-ion

Weight: ≤1.3 lb (0.6 kg)

Charge time (with fully depleted battery): < 190 minutes (typical)

Battery indicators: Each battery has a fuel gauge that indicates its approximate charge. A fuel gauge that shows two or fewer LEDs after a charge cycle indicates that the battery should be replaced.

Charging temperature range: 5° to 45°C (41° to 113°F)

Operating temperature range: 0° to 45°C (32° to 113°F)

Short term (<1 week) storage temperature range: -20° to 60°C (-4° to 140°F)

Long term (>1 week) storage temperature range: 20° to 25°C (68° to 77°F)

Operating and storage humidity range: 5 to 95% relative humidity, non-condensing

For further information, please contact your Stryker representative or visit our website at strykeremergencycare.com

Emergency Care

This document is intended solely for the use of healthcare professionals. A healthcare professional must always rely on his or her own professional clinical judgment when deciding whether to use a particular product when treating a particular patient. Stryker does not dispense medical advice and recommends that healthcare professionals be trained in the use of any particular product before using it.

The information presented is intended to demonstrate Stryker's product offerings. A healthcare professional must always refer to operating instructions for complete directions for use indications, contraindications, warnings, cautions, and potential adverse events, before using any of Stryker's products. Products may not be available in all markets because product availability is subject to the regulatory and/or medical practices in individual markets. Please contact your representative if you have questions about the availability of Stryker's products in your area. Specifications subject to change without notice.

Stryker or its affiliated entities own, use, or have applied for the following trademarks or service marks: CODE SUMMARY, LIFEPAK, LIFEPAK TOUGH, QUIK-COMBO, SpCO, SpMET, Stryker, Masimo, the Radical logo, Rainbow and SET are registered trademarks of Masimo Corporation. All other trademarks are trademarks of their respective owners or holders.

The absence of a product, feature, or service name, or logo from this list does not constitute a waiver of Stryker's trademark or other intellectual property rights concerning that name or logo.




Manufactured by:

Physio-Control, Inc.
11811 Willows Road NE
Redmond, WA, 98052 U.S.A.
Toll free 800 442 1142
strykeremergencycare.com

Stryker European
Operations B.V.

Herikerbergweg 110
1101 CM Amsterdam
Netherlands

 for Physio-Control, Inc.
Tel +31 (0)43 3620008
Fax +31 (0)43 3632001

Distributed by:

Stryker UK Ltd
Stryker House
Hambridge Road
Newbury, Berkshire
RG14 5AW
United Kingdom

Distributed by:

Stryker Australia Pty Ltd
8 Herbert Street
St Leonards NSW 2065
Australia
Toll Free Tel 1800 987 982
Toll Free Fax 1800 890 892